

PATIENT INFORMATION CARD



Ivantis Inc.
201 Technology Dr.
Irvine, CA 92618
USA



+1 949 600 9650



+1 949 600 9651

Place Implant Label Here

Implant Date MM / DD / YYYY OS OD

Surgeon/Physician: _____

This card should be carried by the patient. In the event that an MRI is needed, this card should be given to the health care provider performing the MRI.



MRI Safety Information

Non-clinical testing demonstrated that the Hydrus Microstent is MR Conditional. A patient with this device can be scanned safely in an MR system under the following conditions:

- Static magnetic field of 3-Tesla or less
- Maximum spatial field gradient of 4,000-gauss/cm (40-T/m)
- Maximum MR system reported, whole body averaged specific absorption rate (SAR) of 4-W/kg in the First Level Controlled Operating Mode

Under the scan conditions defined above, the Hydrus Microstent is expected to produce a maximum temperature rise of 2.2°C after 15-minutes of continuous scanning.

In non-clinical testing, the image artifact caused by the Hydrus Microstent extends approximately 2-mm from the Hydrus Microstent when imaged using a gradient echo pulse sequence and a 3-Tesla MR system.